

RayBio[®] Human/Mouse/Rat C-Peptide Enzyme Immunoassay Kit

Catalog #: EIA-CPE, EIAM-CPE, EIAR-CPE

User Manual

Last revised December 1, 2015

Caution:
Extraordinarily useful information enclosed



ISO 13485 Certified

3607 Parkway Lane, Suite 100
Norcross, GA 30092

Tel: 1-888-494-8555 (Toll Free) or 770-729-2992, Fax: 770-206-2393
Web: www.RayBiotech.com, Email: info@raybiotech.com

Table of Contents

Section	Page #
I.	Introduction
II.	General Description
III.	How It Works
IV.	Storage
V.	Reagents
VI.	Additional Materials Required
VII.	Reagent Preparation
	A. Preparation of Plate and Anti-C-Peptide Antibody
	B. Preparation of Biotinylated Peptide (Item F)
	C. Preparation of Standards
	D. Preparation of Positive Control
	E. Preparation of Samples
	F. Preparation of Wash Buffer and HRP-Strep
VIII.	Assay Procedure
IX.	Assay Procedure Summary
X.	Calculation of Results
	A. Typical Data
	B. Sensitivity
	C. Detection Range
	D. Reproducibility
	E. Assay Diagram
XI.	Specificity
XII.	Select Publications
XIII.	Troubleshooting Guide

Please read the entire manual carefully before starting your experiment

I. Introduction

C-Peptide is a 31-amino acid single-chain peptide that is made when proinsulin is split into insulin and C-Peptide. They split before proinsulin is released from endocytic vesicles within the pancreas-one C-Peptide for each insulin molecule. C-Peptide is the middle segment of proinsulin that is between the N-terminal B-chain (30 amino acids) and the C-terminal A-chain (21 amino acids).

Unlike insulin, C-Peptide has no known physiological function. Due to the fact that C-Peptide has a 2-to-5-time longer half-life than insulin, there are higher concentrations of C-Peptide than insulin in the peripheral circulation. Therefore, plasma C-Peptide concentrations may reflect pancreatic insulin secretion more reliably than the level of insulin itself. C-Peptide is cleared from the body by the kidney, and, unlike insulin, urine concentrations are 20 - 50 times higher than in plasma. Unlike plasma insulin levels, which fluctuate in response to meals, measurement of the 24-hour urinary excretion of C-Peptide provides a useful monitor of average beta cell insulin secretion.

The level of C-Peptide has been used to assess the following clinical applications:


- a. to assess the residual beta cell function in patients treated with insulin and to distinguish between types 1 and 2 diabetes. Of particular interest is its use to indicate the need for progression to insulin therapy in type 2 diabetes.
- b. The diagnosis of factitious hypoglycemia. The surreptitious administration of insulin causes high insulin levels in the absence of elevated C-Peptide concentrations.
- c. Insulinoma diagnosis, especially in patients treated with insulin. C-Peptide measurement is used in insulin suppression tests in euglycemic patients with suspected insulinoma. Elevated C-Peptide levels in this test are indicative of insulinoma.
- d. As a marker for residual pancreatic tissue after pancreatectomy. In the case of insulinoma, C-Peptide measurement may be used to detect metastasis and the response to therapy. It may also be used to monitor the progress of pancreas or islet cell transplantation

II. General Description

The RayBio® C-Peptide Enzyme Immunoassay (EIA) Kit is an in vitro quantitative assay for detecting C-Peptide peptide based on the competitive enzyme immunoassay principle.

In this assay, a biotinylated C-Peptide peptide is spiked into the samples and standards. The samples and standards are then added to the plate, where the biotinylated C-Peptide peptide competes with endogenous (unlabeled) C-Peptide for binding to the anti-C-Peptide antibody. After a wash step, any bound biotinylated C-Peptide then interacts with horseradish peroxidase (HRP)-streptavidin, which catalyzes a color development reaction. The intensity of the colorimetric signal is directly proportional to the amount of captured biotinylated C-Peptide peptide and inversely proportional to the amount of endogenous C-Peptide in the standard or samples. A standard curve of known concentration of C-Peptide peptide can be established and the concentration of C-Peptide peptide in the samples can be calculated accordingly.

III. How It Works



IV. Storage

The entire kit may be stored at -20°C to -80°C for up to 6 months from the date of shipment. For extended storage, it is recommended to store at -80°C. **Avoid repeated freeze-thaw cycles.** For prepared reagent storage, see table below.

V. Reagents

Component	Size / Description	Storage / Stability After Preparation
C-Peptide Microplate (Item A)	96 wells (12 strips x 8 wells) coated with secondary antibody.	1 month at 4°C*
Wash Buffer Concentrate (20X) (Item B)	25 ml of 20X concentrated solution.	1 month at 4°C
Standard C-Peptide Peptide (Item C)	2 vials of C-Peptide Peptide. 1 vial is enough to run each standard in duplicate.	The first standard: 2-3 days at 4°C Additional dilutions: Do not store
Anti-C-Peptide Polyclonal Antibody (Item N)	2 vials of anti-C-Peptide.	1 month at 4°C
Assay Diluent A (Item D)	30 ml, contains 0.09% sodium azide as preservative. Diluent for standards and serum or plasma.	N/A
Assay Diluent B (Item E)	15 ml of 5X concentrated buffer. Diluent for standards, cell culture media or other sample types, and HRP-Streptavidin.	1 month at 4°C
Biotinylated C-Peptide Peptide (Item F)	2 vials of Biotinylated C-Peptide Peptide, 1 vial is enough to assay the whole plate.	2-3 days at 4°C
HRP-Streptavidin Concentrate (Item G)	600 µl 200X concentrated HRP-conjugated streptavidin.	Do not store and reuse
Positive Control (Item M)	1 vial of Positive Control.	2-3 days at 4°C
TMB One-Step Substrate Reagent (Item H)	12 ml of 3,3',5,5'-tetramethylbenzidine (TMB) in buffer solution.	N/A
Stop Solution (Item I)	8 ml of 0.2 M sulfuric acid.	N/A

*Return unused wells to the pouch containing desiccant pack, reseal along entire edge.

VI. Additional Materials Required

1. Microplate reader capable of measuring absorbance at 450 nm
2. Precision pipettes to deliver 2 μ l to 1 ml volumes
3. Adjustable 1-25 ml pipettes for reagent preparation
4. 100 ml and 1 liter graduated cylinders
5. Absorbent paper
6. Distilled or deionized water
7. SigmaPlot software (or other software which can perform four-parameter logistic regression models)
8. Tubes to prepare standard or sample dilutions
9. Orbital shaker
10. Aluminum foil
11. Plastic wrap

VII. Reagent Preparation

Keep kit reagents on ice during reagent preparation steps.

Note: **Assay Diluent A** should be used for dilution of samples, Item F and Item C when testing **plasma or serum samples**. **1X Assay Diluent B** should be used for dilution of samples, Item F and Item C when testing **cell culture media or other sample types**.


A. Preparation of Plate and Anti-C-Peptide Antibody

1. Equilibrate plate to room temperature before opening the sealed pouch.
2. Label removable 8-well strips as appropriate for your experiment.
3. 5X Assay Diluent B (Item E) should be diluted 5-fold with deionized or distilled water.
4. Briefly centrifuge the anti-C-Peptide antibody vial (Item N) Then add 50 μ l of 1X Assay Diluent B to the vial to prepare the antibody concentrate. Pipette up and down to mix gently.
5. The antibody concentrate should then be diluted 100-fold with 1X Assay Diluent B. This is your anti-C-Peptide antibody working solution, which will be used in step 2 of Assay Procedure (Section VIII).

Note: The following steps may be done during the antibody incubation procedure (step 2 of Assay Procedure)

B. Preparation of Biotinylated C-Peptide (Item F)

5. Briefly centrifuge the vial of Biotinylated C-Peptide (Item F) before use.
6. See the image below for proper preparation of Item F. Transfer the entire contents of the Item F vial into a tube containing 10 ml of the appropriate Assay Diluent. This is your Working Stock of Item F. Pipette up and down to mix gently. *The final concentration of biotinylated C-Peptide will be **20 ng/ml**.*
 - a. Second Dilution of Item F for Standards: Add 2 ml of Working Stock Item F to 2 ml of the appropriate Assay Diluent. The final concentration of biotinylated C-Peptide will be **10 ng/ml**.
 - b. Second Dilution of Item F for Positive Control: Add 100 μ l of Working Stock Item F to 100 μ l of the prepared Positive Control (Item M). (See section D for Positive Control preparation) The final concentration of biotinylated C-Peptide will be **10 ng/ml**.
 - c. Second Dilution of Item F for samples: Add 125 μ l of Working Stock Item F to 125 μ l of prepared sample (see section E for sample preparation). This is a 2-fold dilution of your sample. The final concentration of biotinylated C-Peptide will be **10 ng/ml**.




C. Preparation of Standards

7. Label 6 microtubes with the following concentrations: 1000 ng/ml, 100 ng/ml, 10 ng/ml, 1 ng/ml, 100 pg/ml and 0 pg/ml. Pipette 450 μ l of biotinylated C-Peptide Item F working solution (prepared in step 6a) into each tube, except the 1,000 ng/ml (leave this one empty).

It is very important to make sure the concentration of biotinylated C-Peptide is 10 ng/ml in all standards.

8. Briefly centrifuge the vial of C-Peptide Standard (Item C). Pipette 8 μ l of Item C and 792 μ l of 10 ng/ml biotinylated C-Peptide working solution (prepared in step 6a) into the tube labeled 1000 ng/ml. Mix thoroughly. This solution serves as the first standard (1,000 ng/ml C-Peptide standard, 10 ng/ml biotinylated C-Peptide).
9. To make the 100 ng/ml standard, pipette 50 μ l of the 1000 ng/ml C-Peptide standard into the tube labeled 100 ng/ml. Mix thoroughly.
10. Repeat this step with each successive concentration, preparing a dilution series as shown in the illustration below. Each time, use 450 μ l of biotinylated C-Peptide and 50 μ l of the prior concentration until the 100 pg/ml is reached. Mix each tube thoroughly before the the next transfer.



D. Positive Control Preparation

11. Briefly centrifuge the Positive Control vial (Item M).
12. Refer to step 6b. This is a 2-fold dilution of the Positive Control. The final concentration of biotinylated C-Peptide should still be 10 ng/ml.

The Positive Control is a cell culture media sample that serves as a system control to verify that the kit components are working. The resulting OD will not be used in any calculations; if no positive competition is observed please contact RayBiotech Technical Support. The Positive Control may be diluted further if desired, but be sure the final concentration of biotinylated C-Peptide is 10 ng/ml.

E. Sample Preparation

13. If you wish to perform a 2-fold dilution of your sample, proceed to step 6c. If you wish to perform a higher dilution of your sample, dilute your sample with the appropriate Assay Diluent before performing step 6c.
EXAMPLE (to make a 4-fold dilution of sample):
 - a. Dilute sample 2-fold (62.5 µl of sample + 62.5 µl of the appropriate Assay Diluent.).
 - b. Perform step 6c (125 µl of working solution Item F + 125 µl of sample prepared above).

The total volume is 250 µl, enough for duplicate wells on the microplate.

It is very important to make sure the final concentration of the biotinylated C-Peptide is **10 ng/ml**.

Note: Optimal sample dilution factors should be determined empirically, however you may reference below for recommended dilution factors for serum: Human=2X
Mouse=2X Rat=2X.

If you have any questions regarding the recommended dilutions you may contact technical support at 888-494-8555 or techsupport@raybiotech.com.

F. Preparation of Wash Buffer and HRP

14. If Item B (20X Wash Concentrate) contains visible crystals, warm to room temperature and mix gently until dissolved.
15. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1X Wash Buffer.
16. Briefly centrifuge the HRP-Streptavidin vial (Item G) before use.
17. Dilute the HRP-Streptavidin concentrate 200-fold with 1X Assay Diluent B.

*Note: do **not** use Assay Diluent A for HRP-Streptavidin preparation in step 17*

VIII. Assay Procedure

1. Keep kit reagents on ice during reagent preparation steps. It is recommended that all standards and samples be run at least in duplicate.
2. Add 100 µl of Anti-C-Peptide Antibody (Item N) (See Reagent Preparation step 3) to each well. Incubate for 1.5 hours at room temperature with gentle shaking (1-2 cycle/sec). You may also incubate overnight at 4°C.
3. Discard the solution and wash wells 4 times with 1X Wash Solution Buffer (200-300 µl each). Washing may be done with a multichannel pipette or an automated plate washer. Complete removal of liquid at each step is essential to good assay performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
4. Add 100 µl of each standard (see Reagent Preparation Section C), Positive Control (see Reagent Preparation Section D) and sample (see Reagent Preparation Section E) in appropriate wells. Be sure to include a blank well (Assay Diluent only). Cover wells and incubate for 2.5 hours at room temperature with gentle shaking (1-2 cycles/sec) overnight or at 4°C.
5. Discard the solution and wash 4 times as directed in Step 3.
6. Add 100 µl of prepared HRP-Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle

shaking. It is recommended that incubation time should not be shorter or longer than 45 minutes.

7. Discard the solution and wash 4 times as directed in Step 3.
8. Add 100 μ l of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking (1-2 cycles/sec).
9. Add 50 μ l of Stop Solution (Item I) to each well. Read at 450 nm immediately.

IX. Assay Procedure Summary

1. Prepare all reagents, samples and standards as instructed.
2. Add 100 μ l anti-C-Peptide to each well. Incubate 1.5 hours at room temperature or overnight at 4°C.
3. Add 100 μ l standard or sample to each well. Incubate 2.5 hours at room temperature or overnight at 4°C.
4. Add 100 μ l prepared Streptavidin solution. Incubate 45 minutes at room temperature.
5. Add 100 μ l TMB One-Step Substrate Reagent to each well. Incubate 30 minutes at room temperature.
6. Add 50 μ l Stop Solution to each well. Read at 450 nm immediately.

X. Calculation of Results

Calculate the mean absorbance for each set of duplicate stands, controls, and samples and subtract the blank optical density. Plot the standard curve using SigmaPlot software (or other software which can perform four-parameter logistic regression models), with standard concentration on the x-axis and percentage of absorbance (see calculation below) on the y-axis. Draw the best-fit curve through the standard points.


Percentage absorbance = $(B - \text{blank OD}) / (B_0 - \text{blank OD})$ where

B = OD of sample or standard and

B_0 = OD of zero standard (total binding)

A. Typical Data

These standard curves are for demonstration only. A standard curve must be run with each assay.



B. Sensitivity

The minimum detectable concentrations of C-Peptide is 772 pg/ml.

C. Detection Range

0.1-1,000 ng/ml

D. Reproducibility

Intra-Assay: CV<10%

Inter-Assay: CV<15%

E. Assay Diagram

Recommended Plate Layout:

Blank	Blank	SA1	SA1	SA9	SA9	SA17	SA17	SA25	SA25	SA33	SA33
Total Binding	Total Binding	SA2	SA2	SA10	SA10	SA18	SA18	SA26	SA26	SA34	SA34
Standard1	Standard1	SA3	SA3	SA11	SA11	SA19	SA19	SA27	SA27	SA35	SA35
Standard2	Standard2	SA4	SA4	SA12	SA12	SA20	SA20	SA28	SA28	SA36	SA36
Standard3	Standard3	SA5	SA5	SA13	SA13	SA21	SA21	SA29	SA29	SA37	SA37
Standard4	Standard4	SA6	SA6	SA14	SA14	SA22	SA22	SA30	SA30	SA38	SA38
Standard5	Standard5	SA7	SA7	SA15	SA15	SA23	SA23	SA31	SA31	SA39	SA39
Pos Control	Pos Control	SA8	SA8	SA16	SA16	SA24	SA24	SA32	SA32	SA40	SA40

Key:

Blank = Buffer Only

Total Binding = Biotin- C-Peptide only

Standard 1 = 1000 ng/ml

Standard 2 = 100 ng/ml

Standard 3 = 10 ng/ml

Standard 4 = 1 ng/ml

Standard 5 = 100 pg/ml

Pos Control = Biotin with Item M

XI. Specificity

This kit only detects C-peptide, not Insulin A or B chain.

Cross Reactivity: This EIA kit shows no cross-reactivity with any of the cytokines tested: Ghrelin, Nesfatin, Angiotensin II, NPY and APC.

XIV. Publications Citing This Product

1. Indumathi S., et al. Nonobese diabetic mice, hypoglycaemia and liver necrosis: a case report. Comparative Clinical Pathology August 2014, released 07 Aug 2014. DOI 10.1007/s00580-014-1969-8
Species: Mouse
Sample Type: Plasma
2. Pujol-Autonell I, Ampudia RM, Monge P, et al. Immunotherapy with Tolerogenic Dendritic Cells Alone or in Combination with Rapamycin Does Not Reverse Diabetes in NOD Mice. ISRN Endocrinol. 2013;2013:346987. <http://dx.doi.org/10.1155/2013/346987>
Species: Mouse
Sample Type: Serum
3. Bhatt MP, Lim Y-C, Hwang J, et al. C-peptide prevents hyperglycemia-induced endothelial apoptosis through inhibition of reactive oxygen species-mediated transglutaminase 2 activation. Diabetes. 2013;62(1):243–253.
Species: Mouse
Sample Type: Serum

XIII. Troubleshooting Guide

Problem	Cause	Solution
Poor standard curve	<ul style="list-style-type: none"> • Inaccurate pipetting • Improper standard dilution 	<ul style="list-style-type: none"> • Check pipettes • Briefly centrifuge Item C and dissolve the powder thoroughly by gently mixing
Low signal	<ul style="list-style-type: none"> • Improper preparation of standard and/or biotinylated antibody • Too brief incubation times • Inadequate reagent volumes or improper dilution 	<ul style="list-style-type: none"> • Briefly spin down vials before opening. Dissolve the powder thoroughly. • Ensure sufficient incubation time; assay procedure step 2 may be done overnight • Check pipettes and ensure correct preparation
Large CV	<ul style="list-style-type: none"> • Inaccurate pipetting • Air bubbles in wells 	<ul style="list-style-type: none"> • Check pipettes • Remove bubbles in wells
High background	<ul style="list-style-type: none"> • Plate is insufficiently washed • Contaminated wash buffer 	<ul style="list-style-type: none"> • Review the manual for proper wash. If using a plate washer, ensure that all ports are unobstructed. • Make fresh wash buffer
Low sensitivity	<ul style="list-style-type: none"> • Improper storage of the ELISA kit • Stop solution 	<ul style="list-style-type: none"> • Follow storage recommendations in sections IV and V. Keep substrate solution protected from light. • Add stop solution to each well before reading plate

RayBio[®] ELISA Kits

Over 2,000 ELISA kits available, visit www.RayBiotech.com/ELISA-Kits.html for details.

This product is for research use only.



©2015 RayBiotech, Inc